



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Oxidized Buffer 3

Part Number: 80-2037-79

Control Number: H26436

Quantity: 2.0 Litres

The above product gave the following results on analysis:

pH = 2.64 @ 20.0 °C

pH Specification = 2.65 ± 0.02 @ 20.0 °C

Conductivity = 15.27 mScm⁻¹ @ 20.0 °C

Conductivity Specification = 15.39 ± 1.22 mScm⁻¹ @ 20.0 °C

Chemical QC

Date: 19 October 2016

Postal Address

Biochrom Ltd
1020 Cambourne Business Park
Cambourne
Cambridge CB23 6DW
England

Telephone

+44 (0)1223 423723

Telefax

+44 (0)1223 420164

Email:

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.