

CERTIFICATE OF ANALYSIS

for

Sodium Oxidised Buffer 4

Part Number: 80-2037-88

Control Number: J26662

Quantity: 2.0 Litres

The above product gave the following results on analysis:

pH = 8.589 @ 20.0 °C

pH Specification = 8.60 ± 0.02 @ 20.0 °C

Conductivity = 38.4 mScm-1 @ 20.0 °C

Conductivity Specification = 38.6 ± 2.76 mScm-1 @ 20.0 °C

Chemical QC

Date: 07 April 2017

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		