



Biochrom Ltd
Certificate No. 890333

CERTIFICATE OF ANALYSIS

for

Sodium Accelerated Buffer 1

Part Number: 80-2115-21

Control Number: J26666

Quantity: 2.0 Litres

The above product gave the following results on analysis:

pH = 3.249 @ 20.0 °C

pH Specification = 3.25 ± 0.02 @ 20.0 °C

Conductivity = 9.6 mScm⁻¹ @ 20.0 °C

Conductivity Specification = 10.0 ± 1.0 mScm⁻¹ @ 20.0 °C

Chemical QC

Date: 10 April 2017

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.