

CERTIFICATE OF ANALYSIS

For

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Batch (Control) Number: J26772

Quantity: 250 mL

Expiry Date: July 2020

The above product gave the following results on analysis:

Baseline Drift:

0.04 $\Delta mV/\Delta t$ at 440nm, 0.003 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Colorimetric Test:

Absorption @ 485nm = 0.039 A

Absorption @ 600nm = 0.001 A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Neil Booth

Date of Issue: 8 August 2017

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.