

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 3

Part Number: 80-2115-23

Batch (Control) Number: J26916

Quantity: 2 Litres

Expiry Date: November 2020

The above product gave the following results on analysis:

pH = 4.386 @ 20.0°C

pH Specification = 4.38 - 4.42 @ 20.0°C

Conductivity = 9.58 mScm⁻¹ @ 20.0°C

Conductivity Specification = 8.0 - 10.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 23 November 2017

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.