

CERTIFICATE OF ANALYSIS

For

Sodium Hydrolysate Buffer 1

Part Number: 80-2037-68

Batch (Control) Number: L27085

Quantity: 2 Litres

Expiry Date: May 2021

The above product gave the following results on analysis:

pH = 3.190 @ 20.0°C

pH Specification = 3.18 - 3.22 @ 20.0°C

Conductivity = 14.94 mScm⁻¹ @ 20.0°C

Conductivity Specification = 14.4 - 16.1 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 1 May 2018

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.