

CERTIFICATE OF ANALYSIS

For

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Batch (Control) Number: L27086

Quantity: 250 mL

Expiry Date: May 2021

The above product gave the following results on analysis:

Baseline Drift:

0.037 $\Delta mV/\Delta t$ at 440nm, 0.008 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Colorimetric Test:

Absorption @ 485nm = 0.043 A

Absorption @ 600nm = 0.001 A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Neil Booth

Date of Issue: 2 May 2018

Postal Address

Telephone

Telefax

Email:

Biochrom Ltd
1020 Cambourne Business Park
Cambourne
Cambridge CB23 6DW
England

+44 (0)1223 423723

+44 (0)1223 420164

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.