

CERTIFICATE OF ANALYSIS

For

Ultra Ninhydrin Solution

Part Number: 80-2117-64

Batch (Control) Number: L27103

Quantity: 250 mL

Expiry Date: June 2021

The above product gave the following results on analysis:

Baseline Drift:

0.025 $\Delta mV/\Delta t$ at 440nm, 0.010 $\Delta mV/\Delta t$ at 570nm
Drift Specification <0.09 at 440nm, <0.06 at 570nm

Colorimetric Test:

Absorption @ 485nm = 0.045 A
Absorption @ 600nm = 0.045 A
Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:
Neil Booth

Date of Issue: 5 June 2018

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.