

CERTIFICATE OF ANALYSIS

For

EZ Nin Reagent NC

Part Number: 80-6003-19

Batch (Control) Number: L27181NC

Quantity: 2 Litres

Expiry Date: 21 August 2021

The above product gave the following results on a standard Lithium High-Performance System:

Baseline Drift:

0.047 $\Delta\text{mV}/\Delta\text{t}$ at 440nm, 0.002 $\Delta\text{mV}/\Delta\text{t}$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Chemical QC Signature:
Daniel Garrod

Date of Issue: 21 August 2018

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.