

# CERTIFICATE OF ANALYSIS

For

## Sodium Oxidized Buffer 4

Part Number: 80-2037-88

Batch (Control) Number: L27339

Quantity: 2 Litres

Expiry Date: December 2021

The above product gave the following results on analysis:

pH = 8.596 @ 20.0°C

pH Specification = 8.58 - 8.62 @ 20.0°C

Conductivity = 38.6 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 35.8 - 41.4 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:  
Neil Booth

Date of Issue: 13 December 2018

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.