

# CERTIFICATE OF ANALYSIS

For

## Ultrosolve Plus

Part Number: 80-2117-65

Batch (Control) Number: M27400

Quantity: 1.75 Litres

Expiry Date: Februray 2022

The above product gave the following results on analysis:

pH = 5.594 @ 20.0°C

pH Specification = 5.55 - 5.75 @ 20.0°C

Conductivity = 18.97 mScm-1 @ 20.0°C

Conductivity Specification = 17.5 - 23.5 mScm-1 @ 20.0°C

Chemical QC Signature:  
Neil Booth

Date of Issue: 6 March 2019

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.