

# CERTIFICATE OF ANALYSIS

For

## Lithium Buffer 1

Part Number: 80-2038-15

Batch (Control) Number: M27475

Quantity: 2 Litres

Expiry Date: June 2022

The above product gave the following results on analysis:

pH = 2.810 @ 20.0°C

pH Specification = 2.78 - 2.82 @ 20.0°C

Conductivity = 14.59 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 14.4 - 16.0 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:  
Daniel Garrod

Date of Issue: 19 June 2019

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.