

# CERTIFICATE OF ANALYSIS

For

## EZ Nin Reagent NC

Part Number: 80-6003-19

Batch (Control) Number: M27585NC

Quantity: 2 Litres

Expiry Date: 23 September 2022

The above product gave the following results on a standard Lithium High-Performance System:

Baseline Drift:

0.048  $\Delta mV/\Delta t$  at 440nm, 0.001  $\Delta mV/\Delta t$  at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Chemical QC Signature:  
Daniel Garrod

Date of Issue: 23 September 2019

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.