

# CERTIFICATE OF ANALYSIS

For

## Lithium Regeneration Buffer 6 NC

Part Number: 80-6003-06

Batch (Control) Number: M27597NC

Quantity: 1 Litre

Expiry Date: October 2022

The above product gave the following results on analysis:

Conductivity = 53.3 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 49.6 - 55.0 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:  
Daniel Garrod

Date of Issue: 16 October 2019

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.