

# CERTIFICATE OF ANALYSIS

For

## Ultra Ninhydrin Solution

Part Number: 80-2117-64

Batch (Control) Number: M27615.

Quantity: 250 mL

Expiry Date: October 2022

The above product gave the following results on analysis:

### Baseline Drift:

0.048.  $\Delta mV/\Delta t$  at 440nm, 0.018.  $\Delta mV/\Delta t$  at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

### Colorimetric Test:

Absorption @ 485nm = 0.040. A

Absorption @ 600nm = 0.008. A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Neil Booth

Date of Issue: 1 November 2019

---

Postal Address

Biochrom Ltd  
1020 Cambourne Business Park  
Cambourne  
Cambridge CB23 6DW  
England

Telephone

+44 (0)1223 423723

Telefax

+44 (0)1223 420164

Email:

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.