

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 3

Part Number: 80-2115-23

Batch (Control) Number: M27637

Quantity: 2 Litres

Expiry Date: December 2022

The above product gave the following results on analysis:

pH = 4.395 @ 20.0°C

pH Specification = 4.38 - 4.42 @ 20.0°C

Conductivity = 9.52 mScm⁻¹ @ 20.0°C

Conductivity Specification = 8.0 - 10.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 27 January 2020

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.