

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 5

Part Number: 80-2037-69

Batch (Control) Number: N27642

Quantity: 2 Litres

Expiry Date: March 2023

The above product gave the following results on analysis:

pH = 3.546 @ 20.0°C

pH Specification = 3.53 - 3.57 @ 20.0°C

Conductivity = 87.7 mScm⁻¹ @ 20.0°C

Conductivity Specification = 82.8 - 89.6 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Daniel Garrod

Date of Issue: 6 March 2020

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.