

CERTIFICATE OF ANALYSIS

For

Sodium Oxidized Buffer 4

Part Number: 80-2037-88

Batch (Control) Number: N27643.

Quantity: 2 Litres

Expiry Date: January 2023

The above product gave the following results on analysis:

pH = 8.593 @ 20.0°C

pH Specification = 8.58 - 8.62 @ 20.0°C

Conductivity = 38.2 mScm⁻¹ @ 20.0°C

Conductivity Specification = 35.8 - 41.4 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 3 February 2020

| | | | |
|--|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.