

# CERTIFICATE OF ANALYSIS

For

## EZ Nin Reagent

Part Number: 80-6000-12

Batch (Control) Number: N27651

Quantity: 2 Litres

Expiry Date: February 2023

The above product gave the following results on a  
standard Lithium High-Performance System:

Baseline Drift:

0.037  $\Delta\text{mV}/\Delta\text{t}$  at 440nm, 0.0099  $\Delta\text{mV}/\Delta\text{t}$  at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Chemical QC Signature:  
Neil Booth

Date of Issue: 20 February 2020

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.