

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 1 NC

Part Number: 80-6003-01

Batch (Control) Number: N27652NC

Quantity: 2 Litres

Expiry Date: 27 February 2023

The above product gave the following results on analysis:

pH = 2.811 @ 20.0°C

pH Specification = 2.78 - 2.82 @ 20.0°C

Conductivity = 15.05 mScm⁻¹ @ 20.0°C

Conductivity Specification = 14.4 - 16.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
David Thomas

Date of Issue: 27 February 2020

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.