

CERTIFICATE OF ANALYSIS

For

Sodium Regeneration Buffer 6

Part Number: 80-2037-57

Batch (Control) Number: N27662

Quantity: 1 Litre

Expiry Date: April 2023

The above product gave the following results on analysis:

Conductivity = 74.9 mScm⁻¹ @ 20.0°C

Conductivity Specification = 68.5 - 75.7 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 28 April 2020

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.