

CERTIFICATE OF ANALYSIS

For

Sodium Loading Buffer

Part Number: 80-2037-67

Batch (Control) Number: N27666

Quantity: 2 Litres

Expiry Date: May 2023

The above product gave the following results on analysis:

pH = 2.197 @ 20.0°C

pH Specification = 2.18 - 2.22 @ 20.0°C

Conductivity = 18.63 mScm⁻¹ @ 20.0°C

Conductivity Specification = 17.7 - 20.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 12 May 2020

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.