

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 4

Part Number: 80-2115-24

Batch (Control) Number: N27706.

Quantity: 2 Litres

Expiry Date: August 2023

The above product gave the following results on analysis:

pH = 9.228 @ 20.0°C

pH Specification = 9.21 - 9.25 @ 20.0°C

Conductivity = 48.5 mScm⁻¹ @ 20.0°C

Conductivity Specification = 47.0 - 51.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:
Neil Booth

Date of Issue: 28 August 2020

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd 1020 Cambourne Business Park Cambourne Cambridge CB23 6DW England	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: 1020 Cambourne Business Park, Cambourne, Cambridge, CB23 6DW England.